

Certificate of Analysis

Product Name : Sigmatrix Urine Diluent
Product Number : SAE0074-50ML
Batch Number : 0000469095
Source Batch : 0000448424
Storage Temperature : +2°C to +8°C
Quality Release Date : 30 Jun 2025

Test	Specification	Result
pH	6.5 - 7.2	7.0
Conductivity (umhos/cm)	16000 - 19000	17580



Jagodige Yasomanee, Supervisor

Quality Assurance

St. Louis, Dekalb

US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchase must determine the suitability of the product for its particular use. See reverse side of website or packing slip for additional terms and conditions of sale

Version Number: 01 Doc: 1231598

Page 1 of 1

